



Challenge

Total protein analysis from small sample volumes in pharmaceutical products.

Solution

Fast, safe and reproducible analysis with a high level of automation and sample throughput using the multi N/C 2100S pharma.

Total Protein Determination in Pharmaceutical Products (Vaccines) by TN_b Analysis on multi N/C 2100S pharma

Introduction

In pharmaceutical vaccine production, starting, intermediate and end products need to be controlled for the level of antigens. They are tested for the quantity of attenuated or devitalized viruses or bacteria. Since these antigens typically consist of proteins, the analytical quantification of total protein is a method of choice. Pharmacopoeia regulations list a number of relevant methods, among them are several UV/Vis assays (e.g. Lowry, Bradford, BCA assay) and two total nitrogen methods, which is the Kjeldahl and the catalytic combustion method. This application note focuses on the catalytic combustion method which is described in Pharm. Eur. Monograph 2.5.33, Method 7 B: high temperature pyrolysis of the nitrogen compounds in an oxygen atmosphere to nitric oxide (NO) followed by chemiluminescence detection (CLD). This method description almost coincides with the EN 12260, which describes TN_b determination in environmental water samples by catalytic high temperature combustion and CLD detection of the formed NO molecules. As this is what multi N/C analyzers are designed for they can be applied for TN_b determination in pharmaceutical products as well. The TN_b can then be converted into total protein to determine the level of antigens.

Since the nitrogen content of proteins varies, it is widely accepted to convert total nitrogen concentrations into total protein concentrations by multiplying a factor of 6.25 according to the following formula:
 $C [\text{Total Protein}] = c [\text{total nitrogen}] \times 6.25$

Materials and Methods

Instrumentation

The analysis was performed on the multi N/C 2100S pharma. The method settings shown in Table 1 were used to determine the TOC content.

Table 1: Method settings

Parameter	multi N/C 2100S pharma
Measurement parameters	TN _b
Digestion	High temperature digestion at 800 °C with platinum catalyst
Number of single repetitions	min. 3, max. 4
Rinse with sample before injection	3 times
Injection volume	75 µL
Dilution	1:1

Samples and Reagents

Five urea TN_b calibration standards prepared at customers' site, one BSA control standard and three unknown customer samples were measured in triplicate determination. The samples were stored at 4 °C in the refrigerator until analysis.

After conditioning at room temperature the liquid samples were directly transferred into 2 mL sample vials using a micro pipet and covered with snap-caps. 75 µL of sample aliquots were transferred into the furnace of the analyzer by aid of the micro liter syringe of the autosampler. Supported by platinum catalyst all nitrogen compounds were converted to nitrogen monoxide (NO) in a pure oxygen atmosphere using catalytic high-temperature combustion. The NO was subsequently detected quantitatively by means of a chemiluminescence detector (CLD).

The measurement sequence was supported by the auto sampler AS 60 with automatic magnetic stirring of the sampling position and syringe wash station.

Calibration

The multi N/C system was calibrated from 5 to 60 mg/L for total bound nitrogen (TN_b) with a multi-point calibration using a BSA protein standard solution. A BSA stock solution of 200 mg TN/L was prepared (Sigma Art.Nr. A-7906, Albumin, Bovine with N-content of 15.60% and purity grade of 98%) weighing 128.2 mg for 100 mL ultra-pure water.

The standard conversion factor for such kind of samples (found in literature) for calculating the protein concentration from the measured N concentration of 6.25 was used to calculate the protein concentration of the measured samples.

The calibration curves can be found in Figure 1.




Figure 1: Example of TN – BSA Calibration curve and characteristics range 5 – 60 mg/L N

Results and Discussion

Each sample has been measured at least three times.

Table 2: Liquid samples

Sample ID	Mean value TN _b [µg/mL] ± RSD [%] (1. vial)	Mean value TN _b [µg/mL] ± RSD [%] (2. vial)	Mean value TN _b [µg/mL] ± RSD [%] (3. vial)
10 ppm Nitrogen Std	10.77 ± 0.14	10.80 ± 0.51	10.74 ± 0.19
20 ppm Nitrogen Std	20.02 ± 0.20	20.05 ± 0.27	19.98 ± 0.24
30 ppm Nitrogen Std	30.31 ± 0.43	30.35 ± 0.28	30.42 ± 0.30
40 ppm Nitrogen Std	40.59 ± 0.27	40.62 ± 0.42	40.65 ± 0.35
50 ppm Nitrogen Std	50.74 ± 0.11	50.96 ± 0.13	50.95 ± 0.29
30 ppm BSA checking Std	30.80 ± 0.54	30.89 ± 0.45	30.88 ± 0.40
Sample A	17.22 ± 0.34	17.20 ± 0.52	17.18 ± 0.44
Sample B	27.95 ± 0.22	27.97 ± 0.21	27.93 ± 0.40
Sample C	48.27 ± 0.34	48.25 ± 0.17	48.05 ± 0.09




Figure 2a: Example 30 ppm BSA checking Std




Figure 2b: Sample A




Figure 2c: Sample B




Figure 2d: Sample C

Conclusion

All samples were measured with very good repeatability and low RSD's.

The clear advantage of a direct injection system using a micro liter syringe is the short and direct sample transfer from the vial to the combustion furnace without long tubings and valves. This allows optimized small rinse volumes of max. three times 75 μ L. Sample consumption can be kept very low. For five replicates including three rinse cycles and a representative injection volume of 75 μ L less than 1.5 mL of sample are required.

multi N/C 2100S pharma equipped with a CLD detector for total nitrogen determination is the optimum system to perform total protein analysis for such samples according to the catalytic high temperature digestion method described in Pharm.Eur. 2.5.33 Method 7B.

The recovery of the customer calibration standard solutions and checking standard (BSA), as well as the presented peak shapes of the sample runs, which show no shifts between replicates nor significant peak tailing, proof the performance of the analyzer.

This document is true and correct at the time of publication; the information within is subject to change. Other documents may supersede this document, including technical modifications and corrections.